

Attestation of Conformity

No. N8AUK 015997 0009 Rev. 00

Holder of Certificate: **Beluk GmbH**
Taubenstr. 1
86956 Schongau
GERMANY

Product: **Electronic measuring equipment**

This Attestation of Conformity is issued on a voluntary basis according to the Electrical Equipment (Safety) Regulations 2016 relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the regulations and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: 713277489

Date, 2023-01-11



(Benedikt Pulver)

Attestation of Conformity

No. N8AUK 015997 0009 Rev. 00

Model(s): CX-Plus, CX-Plus-6T, CX-Plus-12T and CX eco

Parameters:

Voltage measurement:	Rated frequency: 50/60Hz Rated apparent power: 6VA 90-550V 300V LN-519VLL CAT III 550V CAT II
Current measurement:	Rated frequency: 50/60Hz Rated apparent power: 1VA 15mA up to 5A

When installing requirements of Test Standards and Installation Guide must be fulfilled. The equipments are intended for use in a pollution degree 2 or equivalent environment. Requirements for fire enclosure and electrical hazards shall be fulfilled in end-use application

Over current protection is required in the primary and secondary circuit according to Test Standards.

Maximum Surrounding Air Temperature is +60°C.

Tested according to: EN 61010-1:2010/A1:2019
EN IEC 61010-2-030:2021/A11:2021